



**CHHATTISGARH ENVIRONMENT CONSERVATION BOARD**  
**Paryavas Bhawan, North Block, Sector- 19,**  
**Naya Raipur (C.G.)**  
**Email – [hocecb@gmail.com](mailto:hocecb@gmail.com)**

Date: 23/04/2018

**Corrigendum for response of pre-bid queries – Tender No. 05/CECB/e-PROCUREMENT/2017-18**

**Project Name:** "e-TENDER FOR SUPPLY OF SOUND LEVEL METER".

Item	Description
Corrigendum Reference Number	05/CECB/e-PROCUREMENT/2017-18/Corrig1
Tender Reference Number	05/CECB/e-PROCUREMENT/2017-18
Last Date and time for submission of bids	10/05/2018 at 03:00 PM
Physical submission of EMD/DD	10/05/2018 at 03:00 PM
Date of Technical Bid Opening	11/05/2018 at 03:00 PM

Note: Bidder has to submit the signed copy of all the corrigendum with the Technical Bid in envelope A.

## Pre Bid Query Response

No	Section	Page No	RFP Clause	Query	Remarks	Response
1.	4/Technical Bid/Specification for sound level meter- class 2/clause-3	9	Lower limit: Minimum 30 dBA or less than 30 dBA Upper limit: Minimum 130 dBA; in a Single Range.	The Specification should be as 20-140 dBA ; in a Single Range.		No Change. As per RFP.

It is a two bid process. After opening of technical bid demonstration date will be informed by mail after that Commercial bid will be opened online by prior intimation by mail.